

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Listeria Screening PLUS

LOT Lot: 23345

Expiry: 2027-08

REF Art. No.: F5117

specification limit

Control	Cp-value	QC range
Positive Control	28,9	25,6 - 32,2

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	23345	2027-08	o.k.
1	Reaction Mix	23345	2027-08	o.k.
2	Taq Polymerase	23345	2027-08	o.k.
3	Positive Control	23345	2027-08	o.k.

Date 22.08.2025

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.