

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Emetic Bacillus cereus PLUS

LOT Lot: 24175

Expiry: 2027-04

REF Art. No.: F5127

specification limit

Control	Cp-value	QC range
Positive Control	29,1	25,8 - 32,4

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	24175	2027-04	o.k.
1	Reaction Mix	24175	2027-04	o.k.
2	Taq Polymerase	24175	2027-04	o.k.
3	Positive Control	24175	2027-04	o.k.

Date 29.04.2025

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.