

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Campylobacter 4plex

LOT Lot: 25255

Expiry: 2027-06

REF Art. No.: F5170

specification limit

Control	Cp-value	QC range
Positive Control	29,5	26,2 - 32,8

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	25255	2027-06	o.k.
1	Reaction Mix	25255	2027-06	o.k.
2	Taq Polymerase	25255	2027-06	o.k.
3	Positive Control	25255	2027-06	o.k.

Date 30.06.2025

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.