

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Foodborne Pathogens 4plex

LOT Lot: 22215

Expiry: 2027-05

REF Art. No.: F5175

specification limit

Control	Cp-value	QC range
Positive Control	28,8	25,5 - 32,1

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	22215	2027-05	o.k.
1	Reaction Mix	22215	2027-05	o.k.
2	Taq Polymerase	22215	2027-05	o.k.
3	Positive Control	22215	2027-05	o.k.

Date 22.05.2025

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.