

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Enterobacteriaceae 4plex

LOT Lot: 23225

Expiry: 2027-05

REF Art. No.: F5180

specification limit

Control	Cp-value	QC range
Positive Control	29,0	25,7 - 32,3

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	23225	2027-05	o.k.
1	Reaction Mix	23225	2027-05	o.k.
2	Taq Polymerase	23225	2027-05	o.k.
3	Positive Control	23225	2027-05	o.k.

Date 10.06.2025

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.