

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Salmonella ONE

LOT Lot: 24375

Expiry: 2027-09

REF Art. No.: F5211

specification limit

Control	Cp-value	QC range
Positive Control	30,2	26,9 - 33,5

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	24375	2027-09	o.k.
1	Reaction Mix	24375	2027-09	o.k.
2	Taq Polymerase	24375	2027-09	o.k.
3	Positive Control	24375	2027-09	o.k.
L	Lysis Buffer	24375	2027-09	o.k.
L	Lysis Buffer	24375	2027-09	o.k.

Date 22.09.2025

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.