

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Salmonella ONE

LOT Lot: 25054

Expiry: 2026-02

REF Art. No.: F5211

specification limit

Control	Cp-value	QC range
Positive Control	29,4	26,1 - 32,7

Certification

kit code	description	LOT	Expiry	result certification
	Manual	-	-	o.k.
1	Reaction Mix	25054	2026-02	o.k.
1	Reaction Mix	25054	2026-02	o.k.
2	Taq Polymerase	25054	2026-02	o.k.
3	Positive Control	25054	2026-02	o.k.
L	Lysis Buffer	25054	2026-02	o.k.
L	Lysis Buffer	25054	2026-02	o.k.

Date 15.02.2024

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.