

## Quality Assurance Certificate

Version 1.2

Certificate

**Product:** SureFast® Listeria 3plex ONE

**LOT** Lot: 21204

**Expiry:** 2026-05

**REF** Art. No.: F5217

### specification limit

| Control          | Cp-value | QC range    |
|------------------|----------|-------------|
| Positive Control | 30,0     | 26,7 - 33,3 |

### Certification

| kit code | description      | LOT   | Expiry  | result certification |
|----------|------------------|-------|---------|----------------------|
| 1        | Reaction Mix     | 21204 | 2026-05 | o.k.                 |
| 1        | Reaction Mix     | 21204 | 2026-05 | o.k.                 |
| 2        | Taq Polymerase   | 21204 | 2026-05 | o.k.                 |
| 3        | Positive Control | 21204 | 2026-05 | o.k.                 |
| L        | Lysis Buffer     | 21204 | 2026-05 | o.k.                 |
| L        | Lysis Buffer     | 21204 | 2026-05 | o.k.                 |

**Date** 22.05.2024

**Quality Assurance**  
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.