

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® STEC 4plex ONE

LOT Lot: 22424

Expiry: 2026-10

REF Art. No.: F5265

specification limit

Control	Cp-value	QC range
Positive Control	29,5	26,2 - 32,8

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	22424	2026-10	o.k.
1	Reaction Mix	22424	2026-10	o.k.
2	Taq Polymerase	22424	2026-10	o.k.
3	Positive Control	22424	2026-10	o.k.
L	Lysis Buffer	22424	2026-10	o.k.
L	Lysis Buffer	22424	2026-10	o.k.

Date 29.10.2024

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.

