

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Legionella Screen PLUS

LOT Lot: 24174

Expiry: 2026-04

REF Art. No.: F5502

specification limit

Control	Cp-value	QC range
Positive Control	31,7	28,4 - 33,0

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	24174	2026-04	o.k.
1	Reaction Mix	24174	2026-04	o.k.
2	Taq Polymerase	24174	2026-04	o.k.
3	Positive Control	24174	2026-04	o.k.

Date 29.04.2024

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.