

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Legionella Screen PLUS

LOT Lot: 25424

Expiry: 2026-10

REF Art. No.: F5502

specification limit

Control	Cp-value	QC range
Positive Control	32,5	29,2 - 35,8

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	25424	2026-10	o.k.
1	Reaction Mix	25424	2026-10	o.k.
2	Taq Polymerase	25424	2026-10	o.k.
3	Positive Control	25424	2026-10	o.k.

Date 28.10.2024

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.

