

## Quality Assurance Certificate

Version 1.2

Certificate

**Product:** SureFast® Legionella 3plex

**LOT** Lot: 22424

**Expiry:** 2026-10

**REF** Art. No.: F5505

### specification limit

Control	Cp-value	QC range
Positive Control	28,9	25,6 - 32,2

### Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	22424	2026-10	o.k.
1	Reaction Mix	22424	2026-10	o.k.
2	Taq Polymerase	22424	2026-10	o.k.
3	Positive Control	22424	2026-10	o.k.

**Date** 18.10.2024

**Quality Assurance**  
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.