

## Quality Assurance Certificate

Version 1.2

Certificate

**Product:** SureFast® Legionella spp.

**LOT** Lot: 23245

**Expiry:** 2027-06

**REF** Art. No.: F5508

### specification limit

Control	Cp-value	QC range
Positive Control	30,1	26,8 - 33,4

### Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	23245	2027-06	o.k.
2	Taq Polymerase	23245	2027-06	o.k.
3	PCR Water	23245	2027-06	o.k.
3	PCR Water	23245	2027-06	o.k.
4	Positive Control	23245	2027-06	o.k.
4	Positive Control	23245	2027-06	o.k.

**Date** 22.07.2025

**Quality Assurance**  
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.