

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® MRSA 4plex

LOT Lot: 21164

Expiry: 2026-04

REF Art. No.: F7117

specification limit

Control	Cp-value	QC range
Positive Control	25,8	22,5 - 29,1

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	21164	2026-04	o.k.
1	Reaction Mix	21164	2026-04	o.k.
2	Taq Polymerase	21164	2026-04	o.k.
D	Internal Control DNA	21164	2026-04	o.k.
D	Internal Control DNA	21164	2026-04	o.k.
L	Lysis Buffer 1	21164	2026-04	o.k.
L	Lysis Buffer 1	21164	2026-04	o.k.
N	PCR Water	21164	2026-04	o.k.
P	Positive Control	21164	2026-04	o.k.

Date 23.04.2024

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.