

Quality Assurance Certificate

Version 1.2

Certificate

Product: SureFast® Hepatitis E PLUS

LOT Lot: 24375

Expiry: 2027-09

REF Art. No.: F7142

specification limit

Control	Cp-value	QC range
Positive Control	26,6	23,3 - 29,9

Certification

kit code	description	LOT	Expiry	result certification
1	Reaction Mix	24375	2027-09	o.k.
1	Reaction Mix	24375	2027-09	o.k.
2	PP-Mix	24375	2027-09	o.k.
3	Enzyme Mix	24375	2027-09	o.k.
N	PCR Water	24375	2027-09	o.k.
P	Positive Control	24375	2027-09	o.k.
R	Internal Control RNA	24375	2027-09	o.k.
R	Internal Control RNA	24375	2027-09	o.k.

Date 17.09.2025

Quality Assurance
sign. S. Winkelser

Remark: This document is created electronically and therefore valid without signature.